



MAHE INSTITUTE OF DENTAL SCIENCES & HOSPITAL
Affiliated to Pondicherry Central University, Recognized by Dental Council of India
Chalakkara, P.O. Pallor, Mahe-673 310
U.T. of Puducherry. Ph : 0490 2337765

4.5.2: There are established systems and procedures for maintaining and utilizing physical, academic and support facilities - laboratory, library, sports facilities, computers, classrooms etc

INDEX

S.NO	DESCRIPTION	PAGE NUMBER
1	Certificate of the Head of the Institution	2
2	Minutes of the meetings of the Maintenance Committee	4



MAHE INSTITUTE OF DENTAL SCIENCES & HOSPITAL
Affiliated to Pondicherry Central University, Recognized by Dental Council of India
Chalakkara, P.O. Pallor, Mahe-673 310
U.T. of Puducherry. Ph : 0490 2337765

CERTIFICATE OF THE HEAD OF INSTITUTION



MAHE INSTITUTE OF DENTAL SCIENCES & HOSPITAL
Affiliated to Pondicherry Central University, Recognized by Dental Council of India
Chalakkara, P.O. Pallor, Mahe-673 310
U.T. of Puducherry. Ph : 0490 2337765

Dr. ANIL MELATH, MDS.,
PRINCIPAL

TO WHOMSOEVER IT MAY CONCERN

This is to certify that, our Institution has established systems and procedures for maintaining and utilizing physical, academic and support facilities - laboratory, library, sports facilities, computers, classrooms etc. and the details of Minutes of the meetings of Maintenance Committee are given.

PRINCIPAL



MAHE INSTITUTE OF DENTAL SCIENCES & HOSPITAL
Affiliated to Pondicherry Central University, Recognized by Dental Council of India
Chalakkara, P.O. Pallor, Mahe-673 310
U.T. of Puducherry. Ph : 0490 2337765

MAINTENANCE LOG BOOK



MAHE INSTITUTE OF DENTAL SCIENCES & HOSPITAL
Affiliated to Pondicherry Central University, Recognized by Dental Council of India
Chalakkara, P.O. Pallor, Mahe-673 310
U.T. of Puducherry. Ph : 0490 2337765

	Dept. of Public Health Dentistry		
2/7/24	Change two No: Axicator Feet Control	Solved	1 Yes
22/8/24	Change BP-Blower Amp. - B coil	Solved	1 Yes
4/9/24	Create Problems	Solved	1 No
23/9/24	Change tubing	Solved	1 No
23/9/24	Create Problems	Solved	1 No
7/11/24	Keyboard change Cold	Solved	1 Yes
12/12/24	Change Marmust Key Pad.	Solved	1 No
8/1/25	Change Chair Feb and Moving Ke pad	Solved	1 No
9/1/25	Change Coil and Rafter Connector	Solved	1 No
13/1/25	Solenoid valve Problem	Solved	1 No



MAHE INSTITUTE OF DENTAL SCIENCES & HOSPITAL
 Affiliated to Pondicherry Central University, Recognized by Dental Council of India
 Chalakkara, P.O. Pallor, Mahe-673 310
 U.T. of Puducherry. Ph : 0490 2337765

MAHE INSTITUTE OF

SL No.	Department	Complaints	Reported by dept. Incharge	Submitted to
	31/7/24	chair not working		Bio Medical
	2/8/24	Gas leak		Bio Medical
	14/8/24	Acrylizer not working		Bio Medical
	22/8/24	water black spiker comp. block		Bio Man
	24/8/24	water leakage Dental chair		Bio Medical
	29/8/24	Ray faster Glass Broken		Bio Medical
	2/9/24	water leakage Dental chair		Bio Medical
	14/9/24	Dental chair Bulb complaint		Bio Medical
	2/9/24	Door Stopper		Electrition
	24/9/24	Tube not working		Electrition
	24/9/24	Fan not working		Electrition
	27/9/24	Printer Water leakage		Bio medical
	28/9/24	Suction not working		Bio Medical
	17/10/24	Amulien not working		
	10/10/24	Fan not working		Electrition



MAHE INSTITUTE OF DENTAL SCIENCES & HOSPITAL
Affiliated to Pondicherry Central University, Recognized by Dental Council of India
Chalakkara, P.O. Pallor, Mahe-673 310
U.T. of Puducherry. Ph : 0490 2337765

Dept. of Prosthetics

17

Verified by & Date	Service to be done/ Spares be changed	Service done and spares changed	Remarks
9/10/24	Fixing Problem	Solved	1 fan
10/10/24	Change ^{LED} Light PCB	Solved	1 fan
11/10/24	Solved		
10/12/24	Circuit Problem	Solved	1 fan
1/1/25	Water table Sensor	Solved	1 fan
8/1/25	Char PCB Change	Solve	1 fan
8/1/25	"	"	"
23/1/25	Micro motor Contor box ^{After Servicing}		1 fan
23/1/25	changed to fan	solved in	
28/1/25	motherboard issue	Solved.	1 fan
4/3/25	Change Key Pad	Solved	1 fan
5/3/25	Circuit Problem	Solved	1 fan
17/3/25	Chair up-down board Change.	Solved	1 fan
14/4/25	change Auto Release Coil	Solved	1 fan
	monitor problem Replace fan monitor	Solved	1 fan

Dept. of Periodontics

DENTAL SCIENCES & HOSPITAL

18

Verified by & Date	Service to be done/ Spares be changed	Service done and spares changed	Remarks
23/8/24	Change Section tube	Solved	Yes
30/8/24	Change Section Hose	Solved	1 Yes
31/08/24	New monitor installed.	Solved.	cts
8/9/24	Change Spitoon tubing	Solved	Yes
7/9/24	Switch Problems	Solved	1 Yes
9/9/24	Change Low Section tube	Solved	Yes
19/9/24	Change light	solved	
10/9/24	Change Installed Water filter - Moving Sinks	Solved	1 Yes
10/9/24	Installing Water filter	Solved	1 Yes
26/9/24	need to change new bed Pump. High	Solved	cts
07/10/24	motor ISSH sent to service.	Solved 22/10/24	cts
7/10/24	2 No: 6 Secular water problems	Solved	1 Yes
8/10/24	Excretion Filler Prol	Solved	1 Yes
9/10/24	taken for Repair		1 Yes
10/10/24	one Tube extended 3 fan capacitor	Solved	Solved
11/10/24	Change LED tube	Solved	Yes
14/10/24	Secular Unit installed PHD - PERD	Solved	Yes
14/10/24	power ISSH - sent to service.	Solved 22/10/24	cts



MAHE INSTITUTE OF DENTAL SCIENCES & HOSPITAL
 Affiliated to Pondicherry Central University, Recognized by Dental Council of India
 Chalakkara, P.O. Pallor, Mahe-673 310
 U.T. of Puducherry. Ph : 0490 2337765

Dept. of Orthodontics

15

DENTAL SCIENCES & HOSPITAL

Verified by & Date	Service to be done/ Spares be changed	Service done and spares changed	Remarks
8/8/24	Electric Problem	Solved	Nfa
3/8/24	Change 3way Nozzle	Solved	Nfa
25/8/24	Change Water Retractor Valve	Solved	Nfa
24/8/24	Switches Problem	Solved	Nfa
2/9/24	Boat change	Solved	Nfa
4/9/24	Change moving Switches	Solved	Nfa
10-9-24.	filled T connector Water Supply Ok	Solved	Nfa
10/9/24	change moving Switches and Cable	Solved	Nfa
09/9/24	Smps Problem. Temporary install standby Smps. Report to person	Solved.	Stu
23/9/24	— — —	—	Nfa
27/09/24	new Smps installed.	Solved.	Stu
5/10/24	Change Water tube	Solved	Nfa
25/10/24	Change Chair PCB	Solved	Nfa
25/10/24	Change. 3way Spring	Solved	Nfa
26/10/24.	Sent to service.	Solved 04/11/24.	Stu
13/11/24	Change Coil	Solved	Nfa
7/12/24	Retractor Valve Problem	Solved	Nfa



MAHE INSTITUTE OF DENTAL SCIENCES & HOSPITAL
 Affiliated to Pondicherry Central University, Recognized by Dental Council of India
 Chalakkara, P.O. Pallor, Mahe-673 310
 U.T. of Puducherry. Ph : 0490 2337765

MAHE INSTITUTE OF				
Sl No	Department	Reported by dept. incharge	Submitted to	
		CONSERVATIVE DENTISTRY CONSULTANT		
	19/08/24	Pb. RVC system problem.		Stitch
	30/8/24	Chair light problem		Mary
	7/9/24	Chair Motor Complaint		Mary
	18/9/24	Old Complaint		Mary
	24/9/24	Old. Complaint		Mary
	26/9/24	Suction Complaint		Mary
	28/9/24	Afp. Monitor Water problem		
	1/10/24	Chair Cavity Holders Problem		Mary
	4/10/24	Monitor. water block		Mary.
	16/10/24	Suction on/w		Mary
	31/10/24	Old. Problems		Mary
	12/11/24	Recept. 3Way Syringe Leaky		Mary
	12/11/24	Chair motor Cavity problem		Mary
	12/12/24	Suction motor Problems		Mary
	12/12/24	RVC Remote Not working		Mary
	30/12/24	Monitor all w Ph		Mary.

